

Lactanet - Valacta

555, boul. des Anciens Combattants
Ste-Anne-de-Bellevue, QC H9X 3R4
T 514-459-3030 ext 7990
1-800-BON-LAIT
microlab@lactanet.ca



Certificate of Analysis - Somatic Cells Standard Kit

SCC-09-2022

Results are traceable to the certified reference material ERM-BD001 distributed by the European Commission Joint Research Centre

Standard	Concentration 000 cells/ml
low	143
med	520
high	907

For best results:

- * Expiry date: 2022-10-25
- * Do not subdivide samples into smaller portion
- * Warm to $\pm 40^{\circ}\text{C}$ and mix gently to disperse butterfat
- * Discard used vials
- * Do not re-heat samples
- * Do not use leaking vials

Signature 		Date 2022-09-16
--	---	-----------------

Cette version remplace et annule toute version antérieure, le cas échéant. / If applicable, this version replaces all previous issues.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire. / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report