

## Lactanet - Valacta

555, boul. des Anciens Combattants

Ste-Anne-de-Bellevue, QC H9X 3R4

T 514-459-3030 ext 7990

1-800-BON-LAIT

microlab@lactanet.ca



# Certificate of Analysis - Somatic Cells Standard Kit

## SCC-11-2022

Results are traceable to the certified reference material ERM-BD001 distributed by the European Commission Joint Research Centre

Standard	Concentration 000 cells/ml
low	130
med	417
high	883

For best results:

- \* **Expiry date:** 2022-12-20
- \* Do not subdivide samples into smaller portion
- \* Warm to  $\pm 40^{\circ}\text{C}$  and mix gently to disperse butterfat
- \* Discard used vials
- \* Do not re-heat samples
- \* Do not use leaking vials

Signature

Date: 2022-11-14



Cette version remplace et annule toute version antérieure, le cas échéant. / If applicable, this version replaces all previous issues.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire. / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report