

Lactanet - Valacta

555, boul. des Anciens Combattants

Ste-Anne-de-Bellevue, QC H9X 3R4

T 514-459-3030 ext 7990

1-800-BON-LAIT

microlab@lactanet.ca



Certificate of Analysis - Somatic Cells Standard Kit

SCC-03-2024

Results are traceable to the certified reference material ERM-BD001 distributed by the European Commission Joint Research Centre

Standard	Color code	Concentration 000 cells/ml
Low	blue	133
Med	yellow	400
High	red	891

Production date: 2024-03-12

FOR BEST RESULTS

* Expiry date: 2024-04-23

* Do not subdivide samples into smaller portion

* Warm to $\pm 40^{\circ}\text{C}$ and mix gently to disperse butterfat

* Discard used vials

* Do not re-heat samples

* Do not use leaking vials

Signature	Date
	2024-03-14

Cette version remplace et annule toute version antérieure, le cas échéant. / If applicable, this version replaces all previous issues.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire. / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report