

Lactanet - Valacta

555, boul. des Anciens Combattants

Ste-Anne-de-Bellevue, QC H9X 3R4

T 514-459-3030 ext 7990

1-800-BON-LAIT

microlab@lactanet.ca



Certificate of Analysis - Somatic Cells Standard Kit

SCC-05-2024

Results are traceable to the certified reference material ERM-BD001 distributed by the European Commission Joint Research Centre

Standard	Color code	Concentration 000 cells/ml
LOW	blue	131
MED	yellow	471
HIGH	red	860

Production date: 2024-05-14

FOR BEST RESULTS

* Expiry date: 2024-06-25

* Do not subdivide samples into smaller portion

* Warm to $\pm 40^{\circ}\text{C}$ and mix gently to disperse butterfat

* Discard used vials

* Do not re-heat samples

* Do not use leaking vials

Signature	Date 2024-05-17
-----------	------------------------

Cette version remplace et annule toute version antérieure, le cas échéant. / If applicable, this version replaces all previous issues.

Ce certificat ne doit pas être reproduit, sinon en entier, sans l'autorisation écrite du laboratoire. / This certificate shall not be reproduced except in full, without written approval of the laboratory.

Fin du rapport / End of report